

OFFICE OF THE CHIEF MEDICAL OFFICER, CAPITAL HOSPITAL, BHUBANESWAR

Letter No 7881

Date 07th Sep 2012

CLARIFICATION / AMENDMENT IN RESPONSE TO THE PRE BID QUERIES OF THE SPECIFICATION FOR THE TENDER OF MEDICAL EQUIPMENTS

Bid reference-CMO Capital Hospital 2012-2013-ME-001

1. BLOOD GAS ANALYZER:

Under the heading of Operational Requirement the change is **Fully automatic cum electrolyte analyzer instead of Fully automatic , upgradable, fast electrolyte analyzer.**

Under the heading of Eligibility Criteria : The machine should be **USFDA and CE instead of USFDA or CE.**

In technical specification the Essential Measured Parameters pH,pCO₂,pO₂,Barometric Pressure, Na⁺, K⁺ ,Ca⁺⁺ or Cl⁻, Lactic Acid, Hb%, Glucose instead of the Essential Measured Parameters pH, pCO₂, pO₂, Barometric Pressure, Na⁺, K⁺ Ca⁺⁺ , Cl⁻, Lactic Acid, Hb%, Bun Urea, Creatinin, Glucose It should be cartridge based and the expiry interval should be 6-8months.

The Machine should have inbuilt 30 minutes of battery backup.

2. AUTOMATED CELL COUNTER-5PART

The parameters should be **24 as WBC, RBC, HGB, HCT, MCV, MCH, MCHC, PLT, NEUT%, LYMPH%, MONO%, EOS%, BASO%, NEUT#, LYMPH#, MONO#, EOS#, BASO#, RET%, RET#, RDW, PDW, MPV, PCT, IRF instead of 18 parameters.**

GRA the measured parameter should be erased/ deleted.

The measuring principle should be Electrical impedance method or Florescent flow cell method

3. COLOUR DOPPLER ULTRASOUND SCANNER:

No changes and correction required.

4. COLOUR DOPPLER 2D ECHOCARDIOGRAPHY SYSTEM:

Multiplane TEE Probe Should be quoted optionally for Adult and Paediatric. Instead of Multiplane TEE Probe only.

The Terms & conditions of the bid remained unaltered.

Sd/-

Chief Medical Officer, Bhubaneswar

Memo No. 7882

Date

07th Sep 2012

Copy forwarded to System Analyst, Health & F.W department, Government, Odisha, Bhubaneswar for information and necessary action.

He is requested to publish the clarification / amendment in Government website immediately. The soft copy of clarification/ amendment is enclosed herewith for needful.

Chief Medical Officer, Bhubaneswar

Copy to all Notice Board.

10/49